



**Department of Administration / Division of Purchases
One Capitol Hill, Providence, Rhode Island 02908-5855
Web Site: www.purchasing.ri.gov
14 June 11
Addendum # 1**

LOI # 7448710

Title: Data Logger for Vaccine Temperature Monitoring (ARRA Funds)

Submission Deadline: 23 June 11 @ 11:00 AM (Eastern Time)

Correction info:

Page 5: Product Specifications:

B: 2. – Should read Internal Temp sensor (RH is the DNL920 Model, it is different from the DLN910 unit we are requesting and RH is not required for our purposes).

B: 9 - is 59,000 Samples and not 5000

B. DataNet® Wireless Data Logger System Minimum Requirements

1. ZigBee wireless protocol 16-bit data logger
2. Internal RH/Temp sensors
3. 4 Channel RF Logger
4. Fully reliable monitoring network
5. Multiple alarms including Email/Text/SMS to any location
6. Long range monitoring
7. License-free RF band worldwide
8. Continuous temperature monitoring – at least one reading/5 sec
9. Memory: 5000 59,000 readings minimum
10. Download/ archival software package
11. Uncertainty: ± 1 °C (WHO requirement/ industry standard: ± 0.5 °C)
12. Resolution: ± 0.5 °C (industry standard: ± 0.1 °C)
13. Battery life: 6 months
14. Temperature range: -20 °C to 40 °C
15. LED/LCD display with alarm indicators
16. Audible alarm

Jerome D. Moynihan, C.P.M., CPPO
Assistant Director for Special Projects